

# CERTIFICATE OF ACCURACY

Issued by  
Trading Standards



Certificate Number: 20190517/1

Date of Issue: 17<sup>th</sup> May 2019

Page 1 of 2

Trading Standards – Calibration Service  
Flintshire County Council  
County Hall  
Mold  
Flintshire  
CH7 6NB

01352 703181

Email: [compliance@flintshire.gov.uk](mailto:compliance@flintshire.gov.uk)

Web: <http://www.flintshire.gov.uk/>

APPROVED SIGNATORIES  
Rob Crewe  
Inspector of Weights & Measures

## Submitted By

Ash Waste Services

Redwither Road

Wrexham Industrial Estate

Wrexham

LL13 9RD

## Submission Date

n/a

## Date of Test

17<sup>th</sup> May 2019

## DESCRIPTION AND IDENTIFICATION

Make and Model	Bilanciai EV2002
Serial Number	85604
Accuracy Class	III
Max Capacity	50 000 kg
Min Load	400 kg
Scale Interval	20 kg
Location	Ash Waste Services, LL13 9RD

## METHOD OF MEASUREMENT

This is to certify that the relevant class of weights used for the test have been calibrated using standard weights that are traceable to the United Kingdom National Standards via the National Measurement Office (NMO) located Teddington, England, UK.

ALL MAXIMUM PERMISSIBLE ERRORS (MPE) QUOTED BELOW RELATE TO IN SERVICE TOLERANCES

### ACCURACCY

Load (kg)	Indication (kg)	MPE +/- (kg)	
2 000	2 000	20	
4 000	4 000	20	
6 000	6 000	20	
8 000	8 000	20	
10 000	10 000	20	
12 000	12 000	40	
14 000	14 000	40	
16 000	16 000	40	
18 000	18 000	40	
20 000	20 000	40	
31 520	31 520	40	
39 520	39 520	40	
43 800	43 800	60	
<b>PASS</b>	<b>YES</b>	<b>FAIL</b>	<b>N/A</b>

### ECCENTRIC TEST ( )

Position	Indication (g)		
1	5 880		
2	5 880		
3	5 880		
4	5 880		
<b>PASS</b>	<b>YES</b>	<b>FAIL</b>	<b>N/A</b>

### DIAGRAM OF LOAD RECEPTOR



Indicator

### REPEATABILITY TEST

Load	Repeat 1	Repeat 2	Repeat 3	Range	MPE
Not performed	N/A	N/A	N/A	N/A	N/A
<b>PASS</b>	<b>N/A</b>	<b>FAIL</b>	<b>N/A</b>		

END OF CERTIFICATE