

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	Ash Waste
SITE ADDRESS:	Redwither Road
	Wrexam
	North Wales
POST CODE:	LL13 9RD
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	15x3

REPORT NO.:	4879242
INDICATOR TYPE:	WMS 3590ET
INDICATOR SERIAL NO.:	04F514815
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	Wet

REPEATABILITY TEST

Direction →
of
Travel ←

Indicator Reading:	43,100 kg
Weighbridge Viewed from Indicator Position	
Indicator Reading:	43,120 kg

Error Allowed
+/-2dvs or +/-40kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	10,000
--------------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,840	ERROR (e)	10,000	10,000	10,020	0
		error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,020	1.0	Pass		error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass		error +/-2dvs or 40kg
Substitute Load	19,280	19,280	0.0	Pass		error +/-2dvs or 40kg
25,000	24,280	24,240	-2.0	Pass	5,000	error +/-2dvs or 40kg
35,000	34,280	34,260	-1.0	Pass	10,000	error +/-2dvs or 40kg
40,000	39,280	39,280	0.0	Pass	5,000	error +/-2dvs or 40kg
Substitute Load	38,100	38,100	0.0	Pass	0	error +/-2dvs or 40kg
44,000	43,100	43,120	1.0	Pass	5000	error +/-3dvs or 60kg

MAINT LOG UPDATED: NO

MAXIMUM RANGE TESTED: NO

TEST EQUIPMENT USED: Test Weights Traceable to National Standards
27 September 2023

AUTHORISED ENGINEER:
TEST WEIGHT CERTIFICATE NUMBER:
NEXT CALIBRATION DATE:

Steven Fox
Tr11952/Tr11953
25 September 2024

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Name	<i>Lauren Rowley</i>	Date	27/09/2023
Customer Signature	<i>[Signature]</i>	Engineer's Signature	<i>[Signature]</i>

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.