

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	Ash Waste
SITE ADDRESS:	Dunkirk Way
	Cheshire
POST CODE:	CH1 6LZ
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	8x3

REPORT NO.:	4879241
INDICATOR TYPE:	I300
INDICATOR SERIAL NO.:	10F763627
MAXIMUM CAPACITY (kg):	40,000
DIVISION 'e' (kg):	20
Weather Conditions	Wet

REPEATABILITY TEST

Direction →
of
Travel ←

Indicator Reading:	10,000 kg
Weighbridge Viewed from Indicator Position	
Indicator Reading:	10,000 kg

Error Allowed
FALSE
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	15,000
-------------------	--------

		End	Middle	End	Zero
Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	15,000	15,000	15,000	0
3,840	ERROR (e)	error +/-2dvs or 40kg	error +/-2dvs or 40kg	error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass	error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	3,840	3,840	0.0	Pass	error +/-1dvs or 20kg
25,000	8,840	8,840	0.0	Pass	5,000 error +/-1dvs or 20kg
35,000	18,840	18,840	0.0	Pass	10,000 error +/-2dvs or 40kg
40,000	23,840	23,840	0.0	Pass	5,000 error +/-2dvs or 40kg
Substitute Load			0.0	Pass	0 error +/-1dvs or 20kg
44,000			0.0	Pass	5000 error +/-1dvs or 20kg

MAINT LOG UPDATED: NO

MAXIMUM RANGE TESTED: NO

TEST EQUIPMENT USED: Test Weights Traceable to National Standards
CALIBRATION DATE: 27 September 2023

AUTHORISED ENGINEER: Steven Fox
TEST WEIGHT CERTIFICATE NUMBER: Tr11952/Tr11953
NEXT CALIBRATION DATE: 25 September 2024

Comments

BRIDGE TESTED TO 23840 KG WITH AVAILABLE WEIGHTS

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Name	DANIEL NEWLANDS	Date	27/09/2023
Customer Signature	<i>[Signature]</i>	Engineer's Signature	<i>[Signature]</i>

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.