

VERIFICATION CERTIFICATE 7346

CUSTOMER: ALANS SKIP HIRE - ASH SITE ADDRESS: Broughton Industrial Estate Broughton Mills Road Bretton, Chester CH4 0BY MANUFACTURER: Weightron Bilanciai WEIGHBRIDGE TYPE: Steel Surface Mount PLATFORM SIZE: 18m x 3m EST. PLATFORM WEIGHT: 12,000 kg WEIGHBRIDGE LOCATION: Main Weighbridge CUSTOMER EMAIL ADDRESS: melaniemorris@ashgrouppltd.co.uk INITIAL VERIFICATION: <input checked="" type="checkbox"/>	TYPE APPROVAL CERTIFICATE No: LNE-23948 Rev.9 INDICATOR TYPE: X112-A (X112-B i30) INDICATOR TEST CERTIFICATE No: LNE 23273 Rev.6 INDICATOR SERIAL No: 006F1116924 MODULE TEST CERTIFICATE No: N/A MAXIMUM CAPACITY: 50,000 kg DIVISION (e): 20 kg LOADCELL TEST CERTIFICATE No: PTB D09-06.08 PRINTER SERIAL No: N/A TARE FACILITY: Disabled RE-VERIFICATION: <input type="checkbox"/>
---	---

Repeatability Test - MPE 0.3e

	Indication (HR)	Indication (HR)	Indication (HR)
Loaded WTU	43426	43428	43430
	PASS	✓	FAILED

Accuracy of Zero, Linearity/Hysteresis, & Comparison Tests

Approximate Test Interval	MPE	Actual Load	Indication Up (HR)	Error	Corrective Error	Indication Down (HR)	Error	Corrective Error	Comparison
Zero	0.25	0	0	0e		0	0e		
2e	0.5	40	40	0e		40	0e		
20e	0.5	440	440	0e		440	0e		
500e	1.0	10040	10042	0.1e		10040	0e		
1000e	1.0	20040	20042	0.1e					
Substitute Load	1.0	19454	19456	0.1e		19456	0.1e		Yes
1250e	1.0	24454	24456	0.1e		24456	0.1e		
1500e	1.0	29454	29460	0.3e		29460	0.3e		
2000e	1.0	39454	39466	0.6e		39466	0.6e		
Substitute Load	1.0	39418	39430	0.6e		39430	0.6e		
2200e	1.5	43418	43426	0.4e					
PASSED ✓					FAILED				

Loadcell Positions

2	3	4
1	6	5

FLT Substitute Load
3,964 kg

Comparison Test

Printer	
Remote Display	
PC	✓
Other	
Pass	✓
Fail	

Eccentric Load Test – Up to 500e MPE = 0.5e, over 500e MPE = 1.0e

Position	1	2	3	4	5	6	7	8	9	10	11	12
Test Load	10040	10040	10040	10040	10040	10040						
Indication (HR)	10036	10036	10040	10030	10034	10030						
Error	-0.2e	-0.2e	0e	-0.5e	-0.3e	-0.5e						
Corrective Error												
PASSED ✓						FAILED						

ENVIRONMENTAL CONDITIONS: Warm, dry TEST WEIGHTS USED: Precia-Molen UK Ltd VERIFICATION DATE: 16/07/2024	AUTHORISED ENGINEER: Darren Entwistle CERTIFICATE NO: TR 12142 & TR 12143 NEXT CHECK DATE: 16/07/2025
---	--

Signature: Date: ...29/07/2024.....

This Verification Certificate complies with the requirements of Non-Automatic Weighing Instrument Regulations 2016 No: 1152 and EN45501:2015.
 The tests are also carried out in compliance with the UK Weighing Federation Code of Practice for Non Automatic Weighing Instruments.
 Form 080 UK Verification V1.3