



# WEIGHBRIDGE SERVICES

The Chapel, Carrwell Lane, Sheffield, S6 1NR, United Kingdom  
Tel: +44 (0)114 321 6364 / Email: info@weighbridgeservices.com

## CERTIFICATE OF CALIBRATION

<b>Certificate No:</b>	<b>J5445 A</b>	<b>Date:</b>	<b>16/12/24</b>
------------------------	----------------	--------------	-----------------

<b>Company Name:</b>	Ash Waste		
<b>Address:</b>	Ash aggregates LTD		
	Shellway Rd		
	Ellemere port		
	Cheshire		
	CH65 4LQ		
<b>Contact Name:</b>			
<b>Contact Number:</b>			

<b>ID Number:</b>			
<b>Machine Type:</b>	<b>Weighbridge</b>	<b>Manufacturer:</b>	
<b>Model Number:</b>		<b>Serial Number:</b>	<b>1/0742910364</b>
<b>Capacity:</b>	<b>50,000 KG</b>	<b>Increments:</b>	<b>20KG</b>

<b>Owner of Class M1 Test Weights Used:</b>	<b>Crosby</b>
<b>Certificate Number:</b>	
<b>Certificate Number:</b>	

Number of load points (please tick):

<input type="radio"/>	2		3
	END	MIDDLE	END
1			4

<input checked="" type="radio"/>	2	6	3
	END	MIDDLE	END
1	5		4

<input type="radio"/>	2	6	7	3
	END	MIDDLE	END	
1	5	8		4

All Calibrations are performed to UKWF Guidelines

Eccentricity/Corners					
Load Position	Weight Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)
1	10,880	10,860	-20	10,880	0
2	10,880	10,860	-20	10,880	0
3	10,880	10,860	-20	10,880	0
4	10,880	10,860	-20	10,880	0
5	10,880	10,860	-20	10,880	0
6	10,880	10,860	-20	10,880	0
7	N/A	N/A	N/A	N/A	N/A
8	N/A	N/A	N/A	N/A	N/A

  

Load Position	As Found (kg)	Definitive (kg)	Make-up Weights Used: Wagon 19,760kg Forklift 3,880kg
END	N/A	N/A	
MIDDLE	N/A	N/A	
END	N/A	N/A	

Linearity				
Load Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)
0 kg	0 kg	0 kg	0 kg	0 kg
4000 kg	4000 kg	0 kg	4000 kg	0 kg
10,000 kg	9,980 kg	-20 kg	10,000 kg	0 kg
16,000 kg	15,960 kg	-40 kg	16,000 kg	0 kg
20,000 kg	19,960 kg	-40 kg	20,000 kg	0 kg
3,880 kg	3,880 kg	0 kg	3,880 kg	0 kg
19,760 kg	19,720 kg	-40 kg	19,760 kg	0 kg
29,760 kg	29,700 kg	-60 kg	29,760 kg	0 kg
39,760 kg	39,700 kg	-60 kg	39,760 kg	0 kg
43,640 kg	43,580 kg	-60 kg	43,640 kg	0 kg

Repeatability		
Applied Load	As Found (kg)	Definitive (kg)
19,760 kg	19,720 kg	19,760 kg
19,760 kg	19,720 kg	19,760 kg
19,760 kg	19,720 kg	19,760 kg

**Weight and Measures tolerances for use with a Class III Machine:**

**0 - 499 Divisions** = +/- 1 Division  
**500 - 1999 Divisions** = +/- 2 Divisions  
**2000 - 3000 Divisions** = +/- 3 Divisions

<b>Engineer:</b>	<b>Phil Gill</b>
<b>Calibration Date:</b>	<b>6/2/25</b>
<b>Calibration Renewal Date:</b>	<b>Feb 2026</b>



All Calibrations are performed to UKWF Guidelines