



WEIGHBRIDGE SERVICES

The Chapel, Carrwell Lane, Sheffield, S6 1NR, United Kingdom
Tel: +44 (0)114 321 6364 / Email: info@weighbridgeservices.com

CERTIFICATE OF CALIBRATION

Certificate No:	J5445 B	Date:	06/02/25
------------------------	----------------	--------------	-----------------

Company Name:	Ash Waste		
Address:	Ash aggregates LTD		
	Shellway Rd		
	Ellemere port		
	Cheshire		
	CH65 4LQ		
Contact Name:			
Contact Number:			

ID Number:			
Machine Type:	Weighbridge	Manufacturer:	Precia Molen
Model Number:		Serial Number:	64F942564
Capacity:	50,000 KG	Increments:	20KG

Owner of Class M1 Test Weights Used:	Crosby
Certificate Number:	
Certificate Number:	

Number of load points (please tick):

<input type="radio"/>	2		3
	END	MIDDLE	END
1			4

<input checked="" type="radio"/>	2	6	3
	END	MIDDLE	END
1		5	4

<input type="radio"/>	2	6	7	3
	END	MIDDLE	END	
1	5		8	4

All Calibrations are performed to UKWF Guidelines

Eccentricity/Corners					
Load Position	Weight Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)
1	10,880	10,880	0	10,880	0
2	10,880	10,880	0	10,880	0
3	10,880	10,880	0	10,880	0
4	10,880	10,880	0	10,880	0
5	10,880	10,880	0	10,880	0
6	10,880	10,880	0	10,880	0
7	N/A	N/A	N/A	N/A	N/A
8	N/A	N/A	N/A	N/A	N/A

Load Position	As Found (kg)	Definitive (kg)	Make-up Weights Used: Wagon 19,760kg Forklift 3,880kg
END	10,880	10,880	
MIDDLE	10,880	10,880	
END	10,880	10,880	

Linearity				
Load Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)
0 kg	0 kg	0 kg	0 kg	0 kg
4000 kg	4000 kg	0 kg	4000 kg	0 kg
10,000 kg	10,000 kg	0 kg	10,000 kg	0 kg
16,000 kg	16,000 kg	0 kg	16,000 kg	0 kg
20,000 kg	20,000 kg	0 kg	20,000 kg	0 kg
3,880 kg	3,880 kg	0 kg	3,880 kg	0 kg
19,760 kg	19,760 kg	0 kg	19,760 kg	0 kg
29,760 kg	29,760 kg	0 kg	29,760 kg	0 kg
39,760 kg	39,760 kg	0 kg	39,760 kg	0 kg
43,640 kg	43,640 kg	0 kg	43,640 kg	0 kg

Repeatability		
Applied Load	As Found (kg)	Definitive (kg)
19,760 kg	19,760 kg	19,760 kg
19,760 kg	19,760 kg	19,760 kg
19,760 kg	19,760 kg	19,760 kg

Weight and Measures tolerances for use with a Class III Machine:

0 - 499 Divisions = +/- 1 Division
500 - 1999 Divisions = +/- 2 Divisions
2000 - 3000 Divisions = +/- 3 Divisions

Engineer:	Phil Gill
Calibration Date:	6/2/25
Calibration Renewal Date:	Feb 2026



All Calibrations are performed to UKWF Guidelines